




Bulk Asbestos Analysis

By Polarized Light Microscopy
EPA Method: 600/R-93/116 and 600/M4-82-020



Customer: Scientific Analytical Institute, Inc.
4604 Dundas Dr.
Greensboro, NC 27407

Attn: Sharon Donald
Nathan Durham

Lab Order ID: 1407011
Analysis ID: 1407011_PLM
Date Received: 4/22/2014
Date Reported: 4/22/2014

Project: Demonstration PLM Report

Sample ID	Description	Asbestos	Fibrous Components	Non-Fibrous Components	Attributes
Lab Sample ID	Lab Notes				Treatment
Sample - 1 - A	Floor Tile - Kitchen area, beige	None Detected	50% Cellulose 15% Fiber Glass	35% Other	Cream, Orange Fibrous Heterogeneous
1407011PLM_1	floor tile				Teased, Dissolved
Sample - 1 - B	Floor Tile - Kitchen area, beige	8% Chrysotile	10% Cellulose	82% Other	Black Fibrous Heterogeneous
1407011PLM_9	mastic				Dissolved
Sample - 3	Window Glazing	3% Tremolite		97% Other	Gray Non Fibrous Heterogeneous
1407011PLM_3					Ashed, Teased
Sample - 4	Kitchen plaster	None Detected	5% Cellulose 5% Hair	60% Other 30% Quartz	Tan Fibrous Heterogeneous
1407011PLM_4					Teased, Crushed
Sample - 5	Garage - shingle	None Detected	20% Fiber Glass	65% Other 15% Quartz	Black, Tan Fibrous Heterogeneous
1407011PLM_5					Dissolved, Teased
Sample - 6 - A	TSI	5% Chrysotile	10% Fiber Glass 10% Fiber Glass	75% Other	Grayish Non Fibrous Heterogeneous
1407011PLM_6	wrap/mastic				Dissolved
Sample - 6 - B	TSI	5% Amosite	30% Mineral Wool	65% Other	Gray Fibrous Heterogeneous
1407011PLM_10	mud insulation				Teased
Sample - 7	Stucco- Outside Overhang	None Detected		80% Other 20% Quartz	Tan Non Fibrous Heterogeneous
1407011PLM_7					Crushed, Teased

Disclaimer: Due to the nature of the EPA 600 method, asbestos may not be detected in samples containing low levels of asbestos. We strongly recommend that analysis of floor tiles, vermiculite, and/or heterogeneous soil samples be conducted by TEM for confirmation of "None Detected" by PLM. This report relates only to the samples tested and may not be reproduced, except in full, without the written approval of SAL. This report may not be used by the client to claim product endorsement by NVLAP or any other agency of the U.S. government. Analytical uncertainty available upon request. Scientific Analytical Institute participates in the NVLAP Proficiency Testing program. Unless otherwise noted blank sample correction was not performed. Estimated MDL is 0.1%.

Sharon Donald (8)

Analyst

Approved Signatory